



MMC(ASIA)LIMITED

MMC FACTORY TEST CERTIFICATE

工場検査成績書

CERT.No. **G102719**

SHIP'S NAME		HULL No.	
EQUIPMENT	INERT GAS PRESSURE METER		
MODEL	GVP-75	SER. No.	3774
RANGE	VACUUM -10kPa (-1019.72 mmAq) ~ PRESSURE +30kPa (+3059.16 mmAq)		
STANDARD SPECIFICATION OF METER	SPECIFICATION : JIS.B-7505, AU 3/8, ϕ 75 TYPE OF METER : WCA311-231 SER. No. 1404146 BY MIGISHITA SEIKI MFG. CO., LTD		
MANUFACTURER LOCATION	MMC (ASIA) LIMITED 7-7, 2-CHOME, KOTONOO-CHO, CHUO-KU, KOBE 651-0094 JAPAN		

TEST RESULT

INDICATION		ERROR
PRESS(+) or VACUUM(-) on METER	TEST MANOMETER	
+19 kPa (+1937.47 mmAq)	+1908 mmAq	+0.289 kPa (+29.47 mmAq)
+15 kPa (+1529.58 mmAq)	+1516 mmAq	+0.133 kPa (+13.58 mmAq)
+10 kPa (+1019.72 mmAq)	+1027 mmAq	-0.071 kPa (-7.28 mmAq)
+5 kPa (+509.86 mmAq)	+507 mmAq	+0.028 kPa (+2.86 mmAq)
± 0 kPa (ZERO PRESS AND VACUUM)		± 1 kPa area is not applicable to indicate correctly since it is the low pressure
-5 kPa (-509.86 mmAq)	-522 mmAq	-0.119 kPa (-12.14 mmAq)
-10 kPa (-1019.72 mmAq)	-1032 mmAq	-0.120 kPa (-12.28 mmAq)


ROOM TEMPERATURE: 28 °C

ASS'Y INSPECTION



Manufactured by MMC (ASIA) LTD.

FINAL INSPECTION



N. Ohmura / QUALITY CONTROL